

Calibration under low-level of radon activity concentration

Metro
RADON



ELIŠKA FIALOVÁ^{1,2}, PETR OTÁHAL¹, JOSEF VOŠAHLÍK¹

¹National institute for NBC protection (SÚJCHBO, v.v.i.), Kamenná, Czech Republic

²Faculty of Science, Masaryk University in Brno, Czech Republic

September 2021



Introduction




- ❑ Radon measurement techniques are simple, efficient and precise to evaluation the radon activity concentration in indoor or in soil air.
- ❑ Nevertheless the levels of relevant activity concentration in European dwellings are laid down ($300 \text{ Bq}\cdot\text{m}^{-3}$).
- ❑ So the task of developing and accuracy improving new and current calibration procedures for existing commercial radon monitor is still actual.
- ❑ Low-level radon chambre (MetroRadon, $100 - 300 \text{ Bq}\cdot\text{m}^{-3}$)
- ❑ Low-level radon source (TraceRadon, $1 - 100 \text{ Bq}\cdot\text{m}^{-3}$)

Low-level radon chamber

Achieving of low-level radon activity concentration:


- ❑ Constant dosing of radon
- ❑ Defined ventilation
- ❑ Radon free air (specific atmospheric condition in the SUJCHBO areal)
- ❑ Model of constant radon input and constant ventilation

$$a(t) = a_o \cdot e^{-(\lambda+k) \cdot t} + \frac{R}{V(k + \lambda)} (1 - e^{-(\lambda+k) \cdot t})$$



Low-level radon source RF-1 (CMI)

- ❑ Stainless steel cylindrical case, ball valves
- ❑ Steel tray with Ra-226 placed in the middle of this cylindrical case - radon releases from this thin layer
- ❑ Flow-through mode (maximum flow rate 10 l/min)
- ❑ The emanation coefficient was determined by measuring the activity of the RnDP (Pb-214/Bi-214) - the emanation power is almost equal to 1



Český metrologický institut
 Okružní 31, 638 00 Brno
 tel. +420 645 555 111
 www.cmi.cz


Pracoviště: Oblastní inspektorát Praha, Radiová 1136/3, 102 00 Praha 10
 Oddělení výroby standardních radionuklidových zdrojů, Radiová 1288/1a
 tel. +420 266 020 460

CERTIFIKÁT
 etalonu aktivity

Číslo certifikátu: 1035-SE-40175-21	Typ: RF 1	Výrobní číslo: 230321-221458
Radionuklid: Ra-226	Poločas Ra-226: 1600 (7) let	
Aktivita: 1,136 (17) kBq	Poločas Rn-222: 3,8232 (8) dní	
Radonová výdejnost: 0,0023 (1) Bq/s	Emanační schopnost: 0,9552 (19)	
Radionuklidové nečistoty: -	Vnitřní objem mezi ventily: 200 (5) cm ³	
Referenční datum: 1. 4. 2021		




Low-level Rn-222 emanating sources – initial tests for 80 Bq·m⁻³





Low-level Rn-222 emanating sources – initial tests for $40 \text{ Bq}\cdot\text{m}^{-3}$





Calibration laboratory SÚJCHBO



- ❑ Authorized by the Czech Office for Standards, Metrology and Testing
- ❑ Accredited by the Czech Accreditation Institute ($100 \text{ Bq}\cdot\text{m}^{-3} - 2 \text{ MBq}\cdot\text{m}^{-3}$)
- ❑ Certified by Lloyd's Register Quality Assurance





Calibration of radon in soil gas devices in the framework of GARRM 2021



Continual monitors


Grab sampling methods

Please note, that calibration factors mentioned in the new Calibration certificate **should not** be used for results of RIM 2021 - Radon comparison measurement at reference sites Czech Republic!

LOW
3 – 5 kBq·m⁻³

MEDIUM
25 – 40 kBq·m⁻³

HIGH
50 – 70 kBq·m⁻³



15th INTERNATIONAL WORKSHOP GARRM (on the GEOLOGICAL ASPECTS OF RADON RISK MAPPING)

Calibration of radon in soil gas devices in certified laboratory SÚJCHBO

In the framework of the 15th International workshop GARRM it is possible to calibrate radon in soil gas devices in the certified calibration laboratory of SÚJCHBO, v.v.i., Kamenna.

Continual radon monitors will be exposed in the Radon-Aerosol Chamber of SÚJCHBO, v.v.i. under three different levels of radon activity concentration. Time of each exposition will be sent to participants via email briefly after the calibration process.


In the case of grab sampling radon in soil gas devices, participants will take samples of the air from prepared sampling rubber tires with three different levels of radon activity concentration (at most four 150 ml samples from each level of radon activity concentration). It is necessary to participate and operate own device in person in the SÚJCHBO's laboratory.

Based on the participant's results, participants will receive an accredited Calibration protocol.


Calibration laboratory SÚJCHBO v.v.i.

Calibration Laboratory ensures the metrological traceability for devices measuring the radon air concentration and the energy equivalent radon concentration connected with the radon decay products (RaDP). Calibration laboratory is accredited by the Czech Institute for Accreditation according to the standard ČSN EN/IEC 17025:2018 and certified by Lloyd's. Following the metrological traceability, the AMC is on the same metrological level as BfS Berlin. These two institutions compare results periodically and cooperate in some international research projects.

Working plan for calibration



Levels of radon activity concentration



Calibration factors:

Please note, that calibration factors mentioned in the new Calibration certificate **should not** be used for results of RIM 2021 - Radon comparison measurement at reference sites Czech Republic!

Contact persons:

Main person responsible for calibration: Petr Otáhal, Ph.D., otahal@ujchbo.cz, +420 318 600 346
Continual monitor's calibration: Josef Vosabik, vosabik@ujchbo.cz, +420 318 600 209
Grab sampling method's calibration: Eliška Fialová, fialovaeliska@ujchbo.cz, +420 318 600 203





Thanks for your attention!



*The project 19ENV01 TraceRadon has received funding from the EMPIR programme co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme.
19ENV01 traceRadon denotes the EMPIR project reference.*